

SAMPLE

Certificate of Calibration

54924

Product Information

Serial Number:	X06073	Capacity:	500.00 LB
Part Number:	30445-0018	Temperature/Humidity:	70.20 °F/17.90%
Description:	Tension F/G, 500x5LB	Guaranteed Accuracy:	± 1.00% of Capacity
Calibration Date:	2 Mar 06	Uncertainty In Measurement:	4.1557 (unit LB)
Calibration Due:	2 Mar 07		
Calibrated By:	Lonnie Malcolm		
Serial Number:	56E5481		
Unit Condition:	New		

Calibration Equipment Used

Equipment Description:	Dead Weights	Calibration Source:	1100
Equipment Serial Number:	00031	Calibration Date:	12/06/04
Load Range:	750 LB	Calibration Due:	12/06/07
		Calibration Procedure:	26547-0013

Calibration Data

<u>Standard</u> (units LB)	<u>Tolerance</u>		<u>Final From</u> <u>Three Run Average</u> (units LB)	<u>Error</u> (units LB)
	Maximum	Minimum		
100.00	105.00	95.00	104.00	4.00
200.00	205.00	195.00	205.00	5.00
300.00	305.00	295.00	305.00	5.00
400.00	405.00	395.00	398.00	-2.00
500.00	505.00	495.00	495.00	-5.00

Date of Shipment: 03/07/2006

Re-Calibration Date: 1 year after shipment date