

SAMPLE

Certificate of Calibration

55456

Product Information

Serial Number:	X06083	Capacity:	2000.00 LB
Part Number:	XC 30388-0017	Temperature/Humidity:	70.80 °F/27.70%
Description:	X-F/G, COMP., 2000 LB	Guaranteed Accuracy:	± 1.00% of Capacity
Calibration Date:	10 MAR 2006	Uncertainty In Measurement:	9.0702(unit LB)
Calibration Due:	10 MAR 2007		
Calibrated By:	Lonnie Malcolm		
Serial Number:	95F5631		
Unit Condition:	New		

Calibration Equipment Used

Equipment Description:	DTM II (MANUAL)	Calibration Source:	88172
Equipment Serial Number:	0006	Calibration Date:	10/14/05
Load Range:	2000 LB	Calibration Due:	10/14/06
		Calibration Procedure:	26547-0013

Calibration Data

<u>Standard</u> (units LB)	<u>Tolerance</u>		<u>Final From</u> <u>Three Run Average</u> (units LB)	<u>Error</u> (units LB)
	Maximum	Minimum		
400.00	420.00	380.00	388.00	-12.00
800.00	820.00	780.00	782.00	-18.00
1,200.00	1,220.00	1,180.00	1,190.00	-10.00
1,600.00	1,620.00	1,580.00	1,600.67	0.67
2,000.00	2,020.00	1,980.00	2,006.67	6.67

Date of Shipment: 04/03/2006

Re-Calibration Date: 1 year after shipment date