

SAMPLE

CALIBRATION CERTIFICATE

Date:	December 8, 2005	CONDITIONS
Model Number:	DTW 400	Temperature: 74° F
Capacity:	35 lbf-in	Humidity: 15%RH
Accuracy:	0.5% F.S. \pm 1 LSD	
Resolution:	0.01 lbf-in	
Serial Number:	55031	

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

TEST RESULTS

TEST POINT (lbf-in)	ACTUAL (lbf-in)
1) 35.00 (ccw)	35.00
2) 0.0	0.0
3) 35.00 (cw)	35.00
4)	
5)	
6)	
7)	
8)	
9)	
10)	

DATE OF SHIPMENT: December 13, 2005

RECALIBRATION DATE: One year after shipment date