

# SAMPLE

## CALIBRATION CERTIFICATE

**Date:** APRIL 17, 2006  
**Model Number:** ESH-1000N  
**Capacity:** 1000 N  
**Accuracy:** 0.5% F. S.  
**Resolution:** 5 N  
**Serial Number:** 691543

**CONDITIONS**  
Temperature: 74° F  
Humidity: 31%RH

*This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.*

### TEST RESULTS

TEST POINT (N)	ACTUAL (N)
1) 500	503
2) 0	0
3) 1000	1002
4)	
5)	
6)	
7)	
8)	
9)	
10)	

TEST POINT (N)	ACTUAL (N)
11)	
12)	
13)	
14)	
15)	
16)	
17)	
18)	
19)	
20)	

Date of Shipment: 21 APRIL 2006

Re-Calibration Date: 1 year after date of shipment