

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LFS 210		
Serial Number:	269917		
Capacity:	10000	LBS	
Calibration Date:	4/1/2004		
Excitation:	10	VDC	
Calibration Factor:	3.0028	MV/V	Tension
Input Resistance:	375	Ohms	
Output Resistance:	350	Ohms	
Temperature Range:	0-150	Degrees F	

An output of 1.4948 MV/V is induced when a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 4/7/2004 Re-Calibration Date: 1 year after Shipment Date