

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LFS 230		
Serial Number:	1683961		
Capacity:	300	LBS	
Calibration Date:	6/7/2004		
Excitation:	10	VDC	
Calibration Factor:	2.0000	MV/V	Tension
Input Resistance:	532	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	15-104	Degrees F	

An output of **1.4778** **MV/V** is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	GREEN	(+)Excitation
	BLACK	(-) Excitation
	WHITE	(+)Signal
	RED	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 8/6/2004 Re-Calibration Date: 1 year after Shipment Date