

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LFS 250		
Serial Number:	527130-05		
Capacity:	2500	LBS	
Calibration Date:	3/15/2005		
Excitation:	10	VDC	
Calibration Factor:	-3.0022	MV/V	Tension
Input Resistance:	351	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	0-150	Degrees F	

An output of **1.4809** **MV/V** is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	GREEN	(+)Excitation
	BLACK	(-) Excitation
	WHITE	(+)Signal
	RED	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 4/7/2005 Re-Calibration Date: 1 year after Shipment Date