

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LFS 262		
Serial Number:	46493		
Capacity:	500	LBS	
Calibration Date:	10/24/2000		
Excitation:	10	VDC	
Calibration Factor:	2.0493	MV/V	Tension
Input Resistance:	376	Ohms	
Output Resistance:	353	Ohms	
Temperature Range:	60-160	Degrees F	

An output of 1.4969 MV/V is induced when a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 6/14/2005 Re-Calibration Date: 1 year after Shipment Date