

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model: LGP 310
Serial Number: 1022074
Capacity: 50 LBS
Calibration Date: 10/13/2004
Excitation: 10 VDC
Calibration Factor: 3.003 MV/V Tension
Input Resistance: 353 Ohms
Output Resistance: 263 Ohms
Temperature Range: 60-160 Degrees F

An output of 1.120 MV/V is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code

(+)Excitation	A&B
(-) Excitation	C&D
(+)Signal	F
(-) Signal	E

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 2/22/2005 Re-Calibration Date: 1 year after Shipment Date