

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LGP 320		
Serial Number:	147144		
Capacity:	5000	LBS	
Calibration Date:	6/10/2004		
Excitation:	10	VDC	
Calibration Factor:	2.9933	MV/V	Compression
Input Resistance:	350	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	60-160	Degrees F	

An output of **1.4831** **MV/V** is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation	D&F
	BLACK	(-) Excitation	C&E
	GREEN	(+)Signal	A
	WHITE	(-) Signal	B

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 6/17/2004 Re-Calibration Date: 1 year after Shipment Date