

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LGP 380		
Serial Number:	157893		
Capacity:	1000	LBS	
Calibration Date:	12/13/2004		
Excitation:	10	VDC	
Calibration Factor:	2.1328	MV/V	Tension
Input Resistance:	354	Ohms	
Output Resistance:	350	Ohms	
Temperature Range:	15-115	Degrees F	

An output of 1.4547 MV/V is induced when
a shunt resistor of 60K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	(+)Excitation	A
	(-) Excitation	D
	(+)Signal	B
	(-) Signal	C

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 3/23/2005 Re-Calibration Date: 1 year after Shipment Date