

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LGP 382		
Serial Number:	220773		
Capacity:	10000	LBS	
Calibration Date:	5/1/2006		
Excitation:	10	VDC	
Calibration Factor:	-4.1120	MV/V	Compression
Input Resistance:	351	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	15-115	Degrees F	

An output of -2.9078 MV/V is induced when a shunt resistor of 30K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	(+)Excitation	A
	(-) Excitation	D
	(+)Signal	B
	(-) Signal	C

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 6/1/2006 Re-Calibration Date: 1 year after Shipment Date