

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model: LKCP 473-312
Serial Number: 175228
Capacity: 100 LBS
Calibration Date: 2/17/2005
Excitation: 10 VDC
Calibration Factor: 2.0817 MV/V Compression
Input Resistance: 377 Ohms
Output Resistance: 353 Ohms
Temperature Range: 60-160 Degrees F

An output of 1.4784 MV/V is induced when a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code RED (+)Excitation
BLACK (-) Excitation
GREEN (+)Signal
WHITE (-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 2/24/2005 Re-Calibration Date: 1 year after Shipment Date