

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LKCP 472-270	
Serial Number:	88716	
Capacity:	250	LBS
Calibration Date:	5/3/2005	
Excitation:	10	VDC
Calibration Factor:	1.7252	MV/V Compression
Input Resistance:	705	Ohms
Output Resistance:	706	Ohms
Temperature Range:	60-160	Degrees F

An output of 1.7539 MV/V is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 5/5/2005 Re-Calibration Date: 1 year after Shipment Date