

**SAMPLE**

## CERTIFICATE OF CALIBRATION

<b>Transducer Model:</b>	LKCP 476-1.50-MO		
<b>Serial Number:</b>	197370		
<b>Capacity:</b>	3000	LBS	
<b>Calibration Date:</b>	3/10/2006		
<b>Excitation:</b>	10	VDC	
<b>Calibration Factor:</b>	2.1015	MV/V	Compression
<b>Input Resistance:</b>	376	Ohms	
<b>Output Resistance:</b>	351	Ohms	
<b>Temperature Range:</b>	60-160	Degrees F	

An output of 1.4899 MV/V is induced when  
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

### Special Instructions:

<b>Wiring Code</b>	RED	<b>(+)Excitation</b>
	BLACK	<b>(-) Excitation</b>
	GREEN	<b>(+)Signal</b>
	WHITE	<b>(-) Signal</b>

*This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.*

Date of Shipment: 3/14/2006 Re-Calibration Date: 1 year after Shipment Date