

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LKCP 481		
Serial Number:	123890		
Capacity:	130000	LBS	
Calibration Date:	12/19/2002		
Excitation:	10	VDC	
Calibration Factor:	2.0117	MV/V	Compression
Input Resistance:	377	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	60-160	Degrees F	

An output of 1.4844 MV/V is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 12/26/2002 Re-Calibration Date: 1 year after Shipment Date