

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LKCP 484		
Serial Number:	127686		
Capacity:	150000	LBS	
Calibration Date:	10/24/2002		
Excitation:	10	VDC	
Calibration Factor:	2.0383	MV/V	Compression
Input Resistance:	373	Ohms	
Output Resistance:	352	Ohms	
Temperature Range:	60-160	Degrees F	

An output of 1.4879 MV/V is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 10/28/2002 Re-Calibration Date: 1 year after Shipment Date