

**SAMPLE**

## CERTIFICATE OF CALIBRATION

Transducer Model:	LKCP 482-2.13	
Serial Number:	178564	
Capacity:	1000	LBS
Calibration Date:	8/17/2004	
Excitation:	10	VDC
Calibration Factor:	2.3947	MV/V    Compression
Input Resistance:	375	Ohms
Output Resistance:	350	Ohms
Temperature Range:	60-160	Degrees F

An output of 1.4740 MV/V is induced when  
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

### Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

*This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.*

Date of Shipment: 8/20/2004    Re-Calibration Date: 1 year after Shipment Date