

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LKCP 410		
Serial Number:	984642		
Capacity:	20000	LBS	
Calibration Date:	2/26/2004		
Excitation:	10	VDC	
Calibration Factor:	2.279	MV/V	Compression
Input Resistance:	352	Ohms	
Output Resistance:	352	Ohms	
Temperature Range:	60-160	Degrees F	

An output of 1.485 MV/V is induced when a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	WHITE	(+)Signal
	GREEN	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 8/17/2004 Re-Calibration Date: 1 year after Shipment Date