

# SAMPLE

## CERTIFICATE OF CALIBRATION

<b>Transducer Model:</b>	LQB 610		
<b>Serial Number:</b>	1408002901		
<b>Capacity:</b>	1000	LBS	
<b>Calibration Date:</b>	5/25/2010		
<b>Excitation:</b>	10	VDC	
<b>Calibration Factor:</b>	3.000	MV/V	Compression
<b>Input Resistance:</b>	450	Ohms	
<b>Output Resistance:</b>	350	Ohms	
<b>Temperature Range:</b>	14-104	Degrees F	

### Special Instructions:

<b>Wiring Code:</b>	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

*This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.*

Date of Shipment: 5/25/2010 Re-Calibration Date: 1 year after Shipment Date