

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LQB 620		
Serial Number:	224883A		
Capacity:	60	KG	
Calibration Date:	4/7/2003		
Excitation:	10	VDC	
Calibration Factor:	2.1711	MV/V	Tension
Input Resistance:	386	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	60-160	Degrees F	

An output of **1.4774** **MV/V** is induced when
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	GREEN	(+)Excitation
	BLACK	(-) Excitation
	WHITE	(+)Signal
	RED	(-) Signal
	YELLOW	SHIELD

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 6/9/2003 Re-Calibration Date: 1 year after Shipment Date