

## CERTIFICATE OF CALIBRATION

<b>Transducer Model:</b>	LRCN 710		
<b>Serial Number:</b>	45516		
<b>Capacity:</b>	75000	LBS	
<b>Calibration Date:</b>	4/28/2008		
<b>Excitation:</b>	10	VDC	
<b>Calibration Factor:</b>	2.9994	MV/V	Tension
<b>Input Resistance:</b>	350	Ohms	
<b>Output Resistance:</b>	350	Ohms	
<b>Temperature Range:</b>	0-110	Degrees F	

An output of 1.4927 MV/V is induced when  
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

### Special Instructions:

<b>Wiring Code</b>	GREEN	(+)Excitation
	BLACK	(-) Excitation
	WHITE	(+)Signal
	RED	(-) Signal

*This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.*

Date of Shipment: 5/15/2008 Re-Calibration Date: 1 year after Shipment Date