

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LRCN 730		
Serial Number:	38831		
Capacity:	50000	LBS	
Calibration Date:	2/20/2006		
Excitation:	10	VDC	
Calibration Factor:	1.9997	MV/V	Compression
Input Resistance:	450	Ohms	
Output Resistance:	480	Ohms	
Temperature Range:	0-110	Degrees F	

An output of 1.2082 MV/V is induced when
A shunt resistor of 100K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	GREEN	(+)Excitation
	BLACK	(-) Excitation
	WHITE	(+)Signal
	RED	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 7/6/2006 Re-Calibration Date: 1 year after Shipment Date