

## CERTIFICATE OF CALIBRATION

**Transducer Model:** LTL 750-1.25  
**Serial Number:** 101112  
**Capacity:** 25000 LBS  
**Calibration Date:** 4/5/2012  
**Excitation:** 10 VDC  
**Calibration Factor:** 1.6759 MV/V Tension  
**Input Resistance:** 384 Ohms  
**Output Resistance:** 350 Ohms  
**Temperature Range:** -4 to 160 Degrees F

An output of 1.4777 MV/V is induced when  
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

**Special Instructions:** PINS E AND F - N/C

**Wiring Code**

(+)Excitation	D
(-) Excitation	C
(+)Signal	A
(-) Signal	B

*This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.*

Date of Shipment: 5/8/2012

Re-Calibration Date: 1 year after Shipment Date