

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LXT 963		
Serial Number:	178385		
Capacity:	62	IN-LBS	
Calibration Date:	11/8/2004		
Excitation:	10	VDC	
Calibration Factor:	3.1045	MV/V	CW
2nd Calibration Factor:	-3.1003	MV/V	CCW
Input Resistance:	1088	Ohms	
Output Resistance:	1000	Ohms	
Temperature Range:	70-170	Degrees F	

An output of **2.0739** **MV/V** is induced when
a shunt resistor of 120K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	(+)Excitation	A
	(-) Excitation	B
	(+)Signal	C
	(-) Signal	D

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 11/11/2004 Re-Calibration Date: 1 year after Shipment Date