

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LXT 970	
Serial Number:	B150315012	
Capacity:	2.5	NM
Calibration Date:	2/28/2005	
Excitation:	9-12	V
Calibration Factor:	755.95	MV/NM CW
2nd Calibration Factor:	-755.95	MV/NM CCW
Input Resistance:	N/A	Ohms
Output Resistance:	N/A	Ohms
Temperature Range:	32-158	Degrees F

An output of N/A N/A is induced when
N/A

Special Instructions: WHITE=SUPPLY VOLTAGE
BROWN=SIGNAL OUTPUT
BLACK=GROUND
BLUE=NOT USED
GREY=REFERENCE VOLTAGE

Wiring Code	(+)Excitation
	(-) Excitation
	(+)Signal
	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 3/24/2005 Re-Calibration Date: 1 year after Shipment Date