

CERTIFICATE OF CALIBRATION

Transducer Model:	LXT 600		
Serial Number:	XXXXXX		
Capacity:	1200	IN-LBS	
Calibration Date:	2/3/2011		
Excitation:	10	VDC	
Calibration Factor:	2.1360	MV/V	CW
2nd Calibration Factor:	-2.1534	MV/V	CCW
Input Resistance:	746	Ohms	
Output Resistance:	703	Ohms	
Temperature Range:	60-160	Degrees F	

An output of 1.7525 MV/V is induced when

A shunt resistor of 100K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

SAMPLE

Wiring Code:

(+)Excitation A
(-) Excitation B
(+)Signal C
(-) Signal D

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 2/3/2011

Re-Calibration Date: 1 year after Shipment Date