

**SAMPLE**

## CERTIFICATE OF CALIBRATION

Transducer Model:	LXT 912		
Serial Number:	160562		
Capacity:	25	IN-LBS	
Calibration Date:	3/20/2006		
Excitation:	10	VDC	
Calibration Factor:	2.4098	MV/V	CW
2nd Calibration Factor:	-2.4026	MV/V	CCW
Input Resistance:	370	Ohms	
Output Resistance:	351	Ohms	
Temperature Range:	60-160	Degrees F	

An output of **1.4824** **MV/V** is induced when  
a shunt resistor of 59K ohms is applied across (-) Excitation and (-) Signal.

### Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

*This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.*

Date of Shipment: 3/23/2006      Re-Calibration Date: 1 year after Shipment Date