

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	LZBS 1010		
Serial Number:	075751		
Capacity:	65000	LBS	
Calibration Date:	3/15/2004		
Excitation:	5	VDC	
Calibration Factor:	0.9999	MV/V	Compression
Input Resistance:	123	Ohms	
Output Resistance:	123	Ohms	
Temperature Range:	60-160	Degrees F	

An output of 1.695 MV/V is induced when
a shunt resistor of 90K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	RED	(+)Excitation
	BLACK	(-) Excitation
	GREEN	(+)Signal
	WHITE	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 3/18/2004 Re-Calibration Date: 1 year after Shipment Date