

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model: PTT 222-7-F09-15-F2 20K-XKE
Serial Number: 306359
Capacity: 20000 PSI
Calibration Date: 6/18/2003
Excitation: 24 VDC
Calibration Factor: 5.005 VOLTS Pressure
Input Resistance: N/A Ohms
Output Resistance: N/A Ohms
Temperature Range: -20-160 Degrees F

An output of N/A is induced when

Special Instructions: ADDITIONAL WIRING: COMMON=WHITE

Wiring Code	RED	(+)Excitation
	N/A	(-) Excitation
	GREEN	(+)Signal
	N/A	(-) Signal

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 6/19/2003 Re-Calibration Date: 1 year after Shipment Date