

SAMPLE

CERTIFICATE OF CALIBRATION

Transducer Model:	PWD 305	
Serial Number:	1080933	
Capacity:	1	PSI
Calibration Date:	1/13/2006	
Excitation:	10	VDC
Calibration Factor:	1.5023	MV/V Pressure
Input Resistance:	482	Ohms
Output Resistance:	350	Ohms
Temperature Range:	60-160	Degrees F

An output of 0.9750 MV/V is induced when
a shunt resistor of 90K ohms is applied across (-) Excitation and (-) Signal.

Special Instructions:

Wiring Code	(+)Excitation	A&B
	(-) Excitation	C&D
	(+)Signal	F
	(-) Signal	E

This is to certify that the following instrument was calibrated using loading equipment traceable to NIST through one or more of standards. The unit was found to meet or exceed all published sales literature accuracy specifications.

Date of Shipment: 1/18/2006 Re-Calibration Date: 1 year after Shipment Date